Corning Incorporated Life Sciences
---------------------------------------

<u>Registered</u> ISO 9001:2008

# **Product Description**

Height of well

Tolerances of

Dimensions

Opening @ top of well

.989 in.

.143 in. sq.

+/- .010 in.

Catalog Number: 3965

 Product Description:
 Costar® 384-well polypropylene block, standard footprint, non-sterile, bulk

 Component Materials:

 Assay Block

 Virgin Polypropylene, meets USP, Class VI requirements for plastic containers and closures.

# Product Dimensions:Block height-Block length @ bottom-Block width @ bottom-Stance from well center-.177 in.to center

Surface Characterization:

Surface is characterized to be hydrophobic and non-charged.

# **RNase/DNase Testing:**

This product has been tested and is free of any detectable RNase/DNase contamination.

#### Pyrogens:

The product has been tested and has met the criteria established in the *Guideline on Validation of the Limulus Amebocyte Lysate Test as an End-Product Endotoxin Test for Human and Animal Parenteral Drugs, Biological Products and Medical Devices (FDA, December, 1987).* 

## Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:

This product complies with the latest revision of EMEA/410/01 "Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human veterinary medicine products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMEA/410/01.

#### Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures. Visual Attributes: Visual examination of the product. Packaging: Inspection for accurate labeling and correct product configuration.

## Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 8